

QPS Evaluation Services Inc

Testing, Certification and Field Evaluation Body Accredited in Canada, the USA, and Internationally

File LR1356

CERTIFICATE OF COMPLIANCE

(ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to SVANTEK Sp. z o.o

Address Strzyglowska 81

04-872 Warszawa

Poland

Project Number LR1356-2

Product Intrinsically Safe Noise Dosimeter

Model Number SV104BIS

Ratings Ex ia IIC T4 Ga

Class I, Zone 0, AEx ia IIC T4 Ga Ambient temperature: -10°C to +50°C Charging parameters: Um = 8 VDC

Degree of protection: IP65

Applicable Standards CSA C22.2 No. 60079-0:19

UL 60079-0, Edition 7.0

CAN/CSA C22.2 No. 60079-11:14

UL 60079-11, Edition 6.0

Factory/Manufacturing Location SVANTEK Sp. z o.o

Strzyglowska 81, 04-872 Warszawa, Poland

Statement of Compliance: The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Rob Kohuch, Senior Engineer

Signature: Nob Khul- Date: January 6, 2023

QSD 34HL Rev 00